



Config1. (should arrive to last diff. S2D2 with UHV chamber)  $M1=M2=2.13$  mrad, Mono1=20 mm offset, Mono2=15 mm offset  
 Config2. (should arrive to last diff. S2D2 with UHV chamber)  $M1=M2=2.13$  mrad, Mono1=20 mm offset, No Mono2  
 Config3. (should arrive to SiX diff, i.e., first diff in EH2)  $M1=2.13$ ,  $M2=2.72$  mrad, Mono1=20 mm offset, No Mono2  
 Config4. (should arrive to diff in EH1. No matter if the beam does not enter in EH2)  $M1=2.13$ ,  $M2=3.89$  mrad, Mono1=20 mm offset, No Mono2  
 Config5. (should arrive to SiX diff, i.e., first diff in EH2)  $M1=2.72$ ,  $M2=2.72$  mrad, Mono1=20 mm offset, No Mono2  
 Config6. (should arrive to diff in EH1. No matter if the beam does not enter in EH2)  $M1=3.89$ ,  $M2=3.89$  mrad, Mono1=20 mm offset, No Mono2  
 Config7. (should arrive everywhere).  $M1=0$ ,  $M2=0$  mrad, Mono1=20 mm offset, No Mono2  
 Config8. (should arrive everywhere).  $M1=0$ ,  $M2=0$  mrad, Mono1=20 mm offset, Mono2=15 mm offset



**FOLIO 1/4**

Hauteur mediane des airloc  
 Pièces grises: Existantes  
 Pièces rouge: Modifiées, ajout trous  
 Autres couleurs: Nouvelles pièces

ITEM	DRAWING N.	QTY	DESCRIPTION /	REMARKS	MATERIAL	W./Kg
0mm	SCALE:		100	DRN. CKD. APPD.	NAME	DATE
				ISO STANDARD		
				GEN. SURFACE FINISH :		
				GEN. LINEAR TOL. :		
				GEN. ANGULAR TOL. :		
				ASSY.		
						
10			11			12 SolidWorks