




Config1. (should arrive to last diff. S2D2 with UHV chamber) M1=M2=2.13 mrad, Mono1=20 mm offset, Mono2=15 mm offset
 Config2. (should arrive to last diff. S2D2 with UHV chamber) M1=M2=2.13 mrad, Mono1=20 mm offset, No Mono2
 Config3. (should arrive to Si6C diff, i.e., first diff in EH2) M1=2.13, M2=2.72 mrad, Mono1=20 mm offset, No Mono2
 Config4. (should arrive to diff in EH1. No matter if the beam does not enter in EH2) M1=2.13, M2=3.89 mrad, Mono1=20 mm offset, No Mono2
 Config5. (should arrive to Si6C diff, i.e., first diff in EH2) M1=2.72, M2=2.72 mrad, Mono1=20 mm offset, No Mono2
 Config6. (should arrive to diff in EH1. No matter if the beam does not enter in EH2) M1=3.89, M2=3.89 mrad, Mono1=20 mm offset, No Mono2
 Config7. (should arrive everywhere). M1=0, M2=0 mrad, Mono1=20 mm offset, No Mono2
 Config8. (should arrive everywhere). M1=0, M2=0 mrad, Mono1=20 mm offset, Mono2=15 mm offset



FOLIO 2/4

Mono 1 H mini (1410) Mono 2 H maxi (1452)

| ITEM | DRAWING N. | QTY | DESCRIPTION / | REMARKS | MATERIAL | W/Kg |
|---|------------|-----|--|---------|-----------------------|------|
| SCALE: | | | | | NAME | DATE |
| 0mm | | | 100 | | DRN. | |
| | | | | | CKD. | |
| | | | | | APPD. | |
| | | | | | ISO STANDARD | |
| | | | | | GEN. SURFACE FINISH : | |
| | | | | | GEN. LINEAR TOL. : | |
| | | | | | GEN. ANGULAR TOL. : | |
| | | | | | | |
|  | | |  Spanish CRG Beamline AT EUROPEAN SYNCHROTRON BP 220 38043 GRENOBLE CEDEX-FRANCE Comisión Interministerial de Ciencia y Tecnología CICYT C. Rosario Pino 14-16 28020 MADRID - ESPAÑA | | ASSY. | |
| 10 | | | 11 | | 12 SolidWorks | |